



CERTIFICADO DE CONFORMIDAD DE MATERIAL EN CONTACTO CON ALIMENTOS

Alcance: Frasco de PET transparente (**FPET**)
Material: PET

FPET-060-078

FPET-120-187

FPET-250-126

FPET-500-090



Labbox Labware declara, bajo la responsabilidad del fabricante, que los productos indicados en el alcance:

- Reglamento (CE) nº 1935/2004 del Parlamento Europeo y del Consejo, de 27 de octubre de 2004, sobre los materiales y objetos destinados a entrar en contacto con alimentos.
- Cumplen con el reglamento (UE) nº 10/2011 sobre materiales y objetos plásticos destinado a entrar en contacto con los alimentos, el cual determina las condiciones de ensayo de migración a aplicar para cada tipo de muestra.
- El producto, por tanto, no supera el límite de migración total permitido de 10 mg/dm²

06/11/2024

This document has been produced electronically and it is valid without signature.



FOOD CONTACT MATERIAL CONFORMITY CERTIFICATE

Subject: Clear bottle, PET (**FPET**)

Material: PET

FPET-060-078

FPET-120-187

FPET-250-126

FPET-500-090



Labbox labware declare, under the manufacture responsibility, that the product indicated in the subject:

- Comply with Regulation (EC) No 1935/2004 of the European Parliament and of the Council of 27 October 2004 on materials and articles intended to come into contact with food.
- Comply with Regulation (EU) No 10/2011 on plastic materials and articles intended to come into contact with food, which determines the migration test conditions for to be applied in each kind of sample.
- The global migration tests conducted with 10% ethanol and 3% acetic acid do not exceed the permitted total migration limit of 10 mg/dm². These tests were conducted at a temperature of 40°C for a period of 10 days, including heating up to 70°C for a maximum of 2 hours or heating up to 100°C for a maximum of 15 minutes.

06/11/2024

This document has been produced electronically and it is valid without signature.